# **CR300-Series Specifications**



Electrical specifications are valid over a -40 to +70 °C, noncondensing environment, unless otherwise specified. Recalibration is recommended every three years. Critical specifications and system configuration should be confirmed with Campbell Scientific before purchase.

System specifications
Physical specifications
Power requirements 1
Power output specifications 2
Analogue measurement specifications 2
Pulse measurement specifications 3
Digital input/output specifications 4
Communications specifications 4
Standards compliance specifications 5
Warranty 5
Terminal functions

# System specifications

Processor: ARM Cortex M4 running at 144 MHz

#### Memory:

 CPU Drive: 80 MB serial flash • Data Storage: 30 MB serial flash • Operating System: 2 MB flash

• Settings, Calibration, TLS Certificates and Key, System Information: 3 MB serial flash

• Background Tasks, Buffers, System Memory, Table Memory, Program Variables: 756 KB RAM

#### NOTE:

CR300-Series dataloggers with serial numbers 2812 and older have a 5 MB CPU drive and 10 MB serial flash storage. CR300-Series data loggers with serial numbers 2813 and newer, and all CR310 dataloggers have an 80 MB CPU drive and 30 MB serial flash storage.

**Program Execution Period**: 100 ms to 1 day

#### Real-Time Clock:

Battery backed while external power is disconnected

Resolution: 1 ms

• Accuracy: ±1 min. per month

Wiring Panel Temperature: Measured using a thermistor, located on the processor board.

# Physical specifications

**Dimensions** (additional clearance required for cables, wires and antennas):

• **CR300**: 13.97 x 7.62 x 4.56 cm (5.5 x 3.0 x 1.8 in)

• CR310: 16.3 x 8.4 x 5.6 cm (6.4 x 3.3 x 2.2 in)

### Weight/Mass:

• **CR300**: 242 g (0.53 lb)

• CR300-WIFI/RF407/RF412/RF422: 250 g (0.55 lb)

• **CR310**: 288 g (0.64 lb)

• CR310-WIFI/RF407/RF412/RF422: 306 g (0.68 lb)

Case Material: Powder-coated aluminium

# Power requirements

Power specifications for a communications option are shown within the specifications section for that option.

**Protection**: Power inputs are protected against surge, overvoltage, over-current, and reverse power. IEC 61000-4 Class 4 level.

Charger Input (CHG+ and CHG- terminals):

• 16 to 32 VDC

Current limited to 0.9 A maximum

Power converter or solar panel input

**External Batteries** (BAT+ and BAT- terminals):

• 10 to 18 VDC input

• 12 VDC, lead-acid 7 Ah battery, typical

Internal Lithium Battery: 3 V coin cell CR2016 for batterybacked clock. 6-year life with no external power source.

#### Average Current Drain:

Assumes 12 VDC on BAT terminals — add 2 mA if using CHG terminals.

• Idle: 15 mA

• Active 1 Hz scan w/ one analogue measurement: 5 mA

• Active (Processor always on): 23 mA

• **Serial** (RS-232): Active + 25 mA

• Ethernet power requirements (CR310 Only):

• Ethernet idle: 32 mA

• Ethernet link: Active + 51 mA



**USB Power**: Functions that will be active with USB 5 VDC include sending programs, adjusting data logger settings, and making some measurements. If USB is the only power source, then the VX1 and VX2 ranges are reduced to 150 to 2500 mV. The SW12V terminal will not be operational. For the control terminals (C1, C2), voltage output is limited to 4.75 V, and current output is limited to 8 mA.

### Cellular Average Additional Current Contribution at 12 VDC:

- Idle: Connected to network, no data transfer.
  - -CELL200 minimum = 2 mA, average = 10 mA
  - -CELL205 minimum = 2 mA, average = 14 mA
  - -CELL210 minimum = 2 mA, average = 28 mA
  - -CELL215 minimum = 2 mA, average = 14 mA
  - -CELL220 minimum = 2 mA, average = 14 mA
  - -CELL225 minimum = 2 mA, average = 14 mA
- Transfer/Receive:
  - -CELL200 minimum = 20 mA, average = 105 mA
  - -CELL205 minimum = 20 mA, average = 75 mA
  - -CELL210 minimum = 20 mA, average = 90 mA
  - -CELL215 minimum = 20 mA, average = 75 mA
  - -CELL220 minimum = 20 mA, average = 75 mA
  - -CELL225 minimum = 20 mA, average = 75 mA

### Wi-Fi Additional Current Contribution at 12 VDC:

- Client mode communicating: 70 mA typical
- Client mode idle: 7 mA typical
- Access point mode communicating: 70 mA
- Access point mode idle: 62 mA typical
- Idle: <0.1 mA

## RF Average Additional Current Contribution at 12 VDC

	-RF407, -RF412, -RF427	-RF422
Transmit	< 80 mA	20 mA
ldle On	12 mA	9.5 mA
Idle 0.5 s Power Mode	4 mA	3.5 mA
Idle 1 s Power Mode	3 mA	2 mA
Idle 4 s Power Mode	1.5 mA	1.5 mA

# Power output specifications

**System power out limits** (when powered with 12 VDC). Current is limited by a self-resetting thermal fuse.

- 5.8 A @ -40 °C
- 3.7 A @ 20 °C
- 2.0 A @ 70 °C

**VX**: Two independently configurable voltage terminals (VX1-VX2). VX outputs are produced by a 12-bit DAC1. VX terminals can also be used to supply a switched, regulated 5 VDC power source to power digital sensors and toggle control lines.

Range: 150 to 5000 mV
 Resolution: 1.6 mV

 $^{1}\text{Digital}$  to analogue conversion. The process that translates digital voltage levels to analog values.

 Maximum Source Current: 50 mA total, concurrently or independently.

**SW12V**: Provides unregulated 12 VDC power with voltage equal to BAT+ input voltage. SW12V is disabled when operating on USB power only. A thermal fuse regulates current sourcing.

- Thermal Fuse Hold Current (Overload causes voltage drop. Disconnect and let cool to reset. Operate at limit if the application can tolerate some fluctuation.):
  - 1200 mA @ -40 °C
  - 1100 mA @ 20 °C
  - 830 mA @ 70 °C

# Analogue measurement specifications

6 single-ended (SE) or 3 differential (DIFF) terminals individually configurable for voltage, thermocouple, current loop, ratiometric, and period average measurements, using a 24-bit ADC. One channel at a time is measured.

### Voltage measurements

### Terminals:

- Differential Configuration: DIFF 1H/1L 3H/3L
- Single-Ended Configuration: SE1 SE6

### Input Resistance:

- $5 \text{ G}\Omega \text{ typical } (f_{N1} = 50/60 \text{ Hz})$
- 300 M $\Omega$  typical (f<sub>N1</sub> = 4000 Hz)

Input Voltage Limits: -100 to +2500 mV

# Sustained Input Voltage without Damage:

- SE1-SE2: -6 V, +9 V
- SE3-SE6: ±17 V

### DC Common Mode Rejection:

- >120 dB with input reversal
- ≥90 dB without input reversal

### Normal Mode Rejection:

- >71 dB at 50 Hz
- >74 dB at 60 Hz

### Input Current @ 25 °C:

- $\pm .08$  nA typical ( $f_{N1} = 50/60$  Hz)
- $\pm 13 \text{ nA typical } (f_{N1} = 4000 \text{ Hz})$

**Filter First Notch Frequency (f<sub>N1</sub>) Range**: 50/60, 400, 4000 Hz (user specified)

### Analogue Range and Resolution:

			ential Input ersal	Single-Ended and Differential without Input Reversal		
Notch Frequency (f <sub>N1</sub> ) (Hz)	Range <sup>1</sup> (mV)	RMS (µV)	Bits <sup>2</sup>	RMS (µV)	Bits <sup>2</sup>	
4000	-100 to +2500 -34 to +34	23 3.0	16.8 14.5	33 4.2	16.3 14.0	
400	-100 to +2500 -34 to +34	3.8 0.58	19.4 16.8	5.4 0.82	18.9 16.3	
50/60 <sup>3</sup>	-100 to +2500 -34 to +34	1.6 0.23	20.6 18.2	2.3 0.33	20.1 17.7	

<sup>1</sup> Range overhead of ~10% on all ranges guarantees that full-scale values will not cause over range

**Accuracy** (does not include sensor or measurement noise):

• 0 to 40 °C:  $\pm$  (0.04% of measurement + offset)

• -40 to 70 °C:  $\pm (0.1\% \text{ of measurement+ offset})$ 

### Voltage Measurement Accuracy Offsets:

	Typical Offset (μV RMS)						
Range (mV)	Differential with Input Reversal	Differential without Input Reversal	Single- Ended				
-100 to +2500	±20	±40	±60				
-34 to +34	±6	±14	±20				

**Measurement Settling Time**: 20 μs to 600 ms; 500 μs default

### Multiplexed Measurement Time:

Measurement time = (multiplexed measurement time + settling time) • reps +0.8 ms

	Differential with Input Reversal	Single-Ended or Differential without Input Reversal
Example fN1 <sup>1</sup> (Hz)	Time <sup>2</sup> (ms)	Time <sup>2</sup> (ms)
4000	2.9	1.4
400	14.6	7.3

	Differential with Input Reversal	Single-Ended or Differential without Input Reversal
Example fN1 <sup>1</sup> (Hz)	Time <sup>2</sup> (ms)	Time <sup>2</sup> (ms)
50/60	103	51.5

<sup>&</sup>lt;sup>1</sup> Notch frequency (1/integration time).

## Resistance measurement specifications

The data logger makes ratiometric-resistance measurements for four- and six-wire full-bridge circuits and two-, three-, and four-wire half-bridge circuits using voltage excitation.

### Accuracy:

Assumes input reversal for differential measurements **RevDiff**. Does not include bridge resistor errors or sensor and measurement noise.

- 0 to 40 °C:  $\pm$ (0.05% of voltage measurement + offset)
- -40 to 70 °C:  $\pm$ (0.06% of voltage measurement + offset)

# Current-loop measurement specifications

Two analogue inputs terminals may be configured as independent, non-isolated 0-20 mA or 4-to-20 mA current-loop inputs referenced to ground. One channel at a time is measured. Current is measured using a 24-bit ADC<sup>1</sup>.

**Terminals**: SE1-SE2 **Range**: 0 to 25 mA

### Accuracy:

0 to 40 °C: ±0.14% of reading
-40 to 70 °C: ±0.26% of reading

# Pulse measurement specifications

Terminals are individually configurable for switch closure, high-frequency pulse, or low-level AC measurements.

### Switch-closure input

### Terminals:

P SW

• C1-C2 (Requires an external 100 k $\Omega$  resistor connected from the terminal to VX1 or VX2.)

Maximum Input Frequency: 150 Hz Minimum Switch Closed Time: 3 ms Minimum Switch Open Time: 3 ms

Maximum Bounce Time: 1 ms open without being counted

<sup>&</sup>lt;sup>2</sup> Typical effective resolution (ER) in bits; computed from ratio of full-scale range to RMS resolution.

 $<sup>^{\</sup>rm 3}$  50/60 corresponds to rejection of 50 and 60 Hz ac power mains noise.

<sup>&</sup>lt;sup>2</sup> Default settling time of 500 µs used.

<sup>&</sup>lt;sup>1</sup>Analogue to digital conversion. The process that translates analogue voltage levels to digital values.

# High-frequency input

#### Terminals:

• SE1-SE4

P LL

P SW

C1-C2

### Maximum Input Frequency:

SE1-SE4: 35 kHzP\_LL: 20 kHzP\_SW: 35 kHzC1-C2: 3 kHz

## Low-level AC input

Terminal: P\_LL

Maximum Input Voltage: ±20 VDC

DC-offset Rejection: Internal AC coupling eliminates DC-offset

voltages up to  $\pm 0.05$  VDC

Input Hysteresis: 12 mV at 1 Hz Low-Level AC Pulse Input Ranges:

Sine wave (mV RMS)	Range (Hz)
20	1.0 to 20
200	0.5 to 200
2000	0.3 to 10,000
5000	0.3 to 20,000

# Quadrature input

**Terminals**: SE1 and SE2 or C1 and C2 can be configured as digital terminal pairs to monitor the two sensing channels of an encoder.

Maximum Frequency: 2.5 kHz

# Period-averaging measurement specifications

Terminals: SE1-SE4

**Accuracy**:  $\pm$  (0.01% of measurement + resolution), where resolution is 0.13  $\mu$ s divided by the number of cycles to be

measured

**Voltage Range**: 0 to 3.3 V **Minimum Pulse Width**: 2.5 μs

Voltage Threshold: Counts cycles on transition from < 0.9 VDC

to >2.1 VDC

# Digital input/output specifications

Up to seven terminals may be configured for digital input or output (I/O).

#### Terminals:

• SE1-SE4

P SW

• C1-C2

### Digital I/O Voltage Levels:

Terminal	High State	Low State	Current Source	Maximum Input Voltage
C1 C2	5.0 V output 3.3V input	0 V	10 mA at 3.5 V	–10 V, +15 V
SE1 SE2	3.3 V	0 V	100 μA at 3.0 V	−6 V, +9 V
SE3 SE4 P_SW	3.3 V	0 V	100 μA at 3.0 V	±17 V

### Pulse-width modulation

### Terminals<sup>1</sup>

SE1-SE4

Period Maximum: 2047 ms

#### Resolution

• 0 – 5 ms: 83.33 ns or 12 MHz

• 5 – 325 ms: 5.00 µs or 200 kHz

• > 325 ms: 31.25 µs or 32 kHz

# Communications specifications

**Ethernet Port** (CR310 Only): RJ45 jack, 10/100Base Mbps, full and half duplex, Auto-MDIX, magnetic isolation, and TVS surge protection.

Internet Protocols: Ethernet, PPP, RNDIS, ICMP/Ping, Auto-IP (APIPA), IPv4, IPv6, UDP, TCP, TLS (v1.2), DNS, DHCP, SLAAC, Telnet, HTTP(S), FTP(S), POP3/TLS, NTP, SMTP/TLS

**Additional Protocols**: PakBus, PakBus Encryption, SDI-12, Modbus RTU / ASCII / TCP, DNP3, custom user definable over serial, UDP

**USB Device**: Micro-B device for computer connectivity

**SDI-12** (C1, C2): Two independent SDI-12 compliant terminals are individually configured and meet SDI-12 Standard v 1.4.

**RS-232**: Female RS-232, 9-pin interface, 1200 to 115.2 kbps

# Cellular option specifications

### Cell Technology:

Option	Cellular Protocol
-CELL200	3G, 2G
-CELL205	4G LTE with automatic 3G fallback
-CELL210	4G LTE CAT-1
-CELL215	4G LTE with automatic 3G fallback
-CELL220	4G LTE with automatic 3G fallback
-CELL225	4G LTE

#### See

https://s.campbellsci.com/documents/us/miscellaneous/Cellular%20Modem%20Frequency%20Bands.pdf for a complete list of supported frequency bands.

Antenna Terminal: SMA

**SIM Slot**: Industry standard 3FF micro-SIM (6 position / contacts)

# Wi-Fi option specifications

WLAN (Wi-Fi) (CR300-WIFI only)

Maximum Possible Over-the-Air Data Rates: <11 Mbps over 802.11b, <54 Mbps over 802.11g, <72 Mbps over 802.11n

**Operating Frequency**: 2.4 GHz, 20 MHz bandwidth **Antenna Connector**: Reverse Polarity SMA (RPSMA)

Antenna (shipped with data logger): Unity gain (0 dBd), 1/2 wave whip, omnidirectional. Features an articulating knuckle joint that can be oriented vertically or at right angles

**Supported Technologies**: 802.11 b/g/n, WPA/WPA2-Personal, WPA AM/PA2 Enterprise Security WEB

WPA/WPA2-Enterprise Security, WEP

Client Mode: WPA/WPA2-Personal and Enterprise, WEP

Access Point Mode: WPA2-Personal

Receive Sensitivity: -97 dBm

# RF radio option specifications

**Antenna Terminal**: Reverse Polarity SMA (RPSMA)

# Radio Type

- **RF407**, **RF412**, and **RF427**: Frequency Hopping Spread Spectrum (FHSS)
- **RF422**: SRD860 Radio with Listen before Talk (LBT) and Automatic Frequency Agility (AFA)

### Frequency

- **RF407**: 902 to 928 MHz (US, Canada)
- **RF412**: 915 to 928 MHz (Australia, New Zealand)
- **RF422**: 863 to 870 MHz (European Union)
- RF427: 902 to 907.5 MHz/915 to 928 MHz (Brazil)

### **Transmit Power Output** (software selectable)

- RF407 and RF412: 5 to 250 mW
- RF422: 2 to 25 mW
- RF427: 5 to 250 mW

### **Channel Capacity**

- **RF407**: Eight 25-channel hop sequences sharing 64 available channels.
- **RF412**: Eight 25-channel hop sequences sharing 31 available channels.
- RF422: Ten 30-channel hop sequences (default), software configurable to meet local regulations; 10 sequences for reducing interference through channel hop.
- **RF427**: Eight 25-channel hop sequences sharing 43 available channels.

### Receive Sensitivity

- RF407, RF412, and RF427: -101 dBm
- **RF422**: -106 dBm

#### RF Data Rate

- RF407, RF412, and RF427: 200 kbps
- **RF422**: 10 kbps

# Standards compliance specifications

View EU Declarations of Conformity at www.campbellsci.eu/cr300 and www.campbellsci.eu/cr310.

Shock and Vibration: ASTM D4169-09

Protection: IP30

### EMI and ESD protection:

- Immunity: Meets or exceeds following standards:
  - **ESD**: per IEC 61000-4-2; ±15 kV air, ±8 kV contact discharge
  - Radiated RF: per IEC 61000-4-3; 10 V/m, 80-1000 MHz
  - EFT: per IEC 61000-4-4; 4 kV power, 4 kV I/O
  - Surge: per IEC 61000-4-5; 4 kV power, 4kV I/O
  - Conducted RF: per IEC 61000-4-6; 10 V power, 10
     V I/O
- Emissions and immunity performance criteria available on request.

### RF407 Option

- United States FCC Part 15.247: MCQ-XB900HP
- Industry Canada (IC): 1846A-XB900HP
- Mexico IF: RCPDIXB15-0672-A1

### RF412 Option

- ACMA RCM
- United States FCC Part 15.247:
- MCO-XB900HP
- Industry Canada (IC): 1846A-XB900HP

**RF422 Option**: View EU Declaration of Conformity at www.campbellsci.eu/cr300 and www.campbellsci.eu/cr310.

**RF427 Option**: Brazil ANATEL standards in Resolution No. 506: 08335-17-10644. View the RF427 Brazilian Certificate of Conformity at www.campbellsci.eu/cr300 and www.campbellsci.eu/cr310.

### WIFI Option

- United States FCC ID: XF6-RS9113SB
- Industry Canada (IC): 8407A-RS9113SB

## Cellular Option:

• Industry Canada (IC): 10224A-201611EC21A

#### NOTE

The user is responsible for emissions if changing the antenna type or increasing the gain.

# Warranty

**Standard**: Three years against defects in materials and workmanship.

# Terminal functions

Analogue input terminal functions								
SE DIFF	3 4 <sub>「</sub> 2 <sub>7</sub> H L		5 6 <sup>73</sup> 7 H L					
Single-Ended Voltage	✓	✓	✓	✓	✓	✓		
Differential Voltage	Н	L	Н	L	Н	L		
Ratiometric/Bridge	✓	✓	✓	✓	✓	✓		
Thermocouple	✓	✓	✓	✓	✓	✓		
Current Loop	✓	✓						

Pulse counting terminal functions										
Pulse Counting	C1	C2	P_ SW	P_ LL	SE1	SE2	SE3	SE4	SE5	SE6
Switch- Closure	✓	<b>✓</b>	<b>✓</b>							
High Frequency	✓	<b>✓</b>								
Low-level AC				<b>✓</b>						
Quadrature	✓	✓			✓	✓				
Period Average					<b>✓</b>	<b>✓</b>	<b>✓</b>	<b>✓</b>		

Analog <b>ue</b> output terminal functions							
VX1 VX2							
Switched Voltage Excitation							

Voltage output terminal functions							
	<b>C</b> 1	C2	SE1-4	VX1	VX2	P_SW	SW12V
3.3 VDC			✓	✓	✓	✓	
5 VDC	✓	✓		✓	✓		
BAT +							✓

Communications terminal functions							
	<b>C</b> 1	C2	SE1-3	RS-232			
SDI-12	<b>✓</b>	<b>✓</b>					
RS-232				✓			
RS-232 0-5V	✓	✓					
GPS Time Sync	✓	✓	✓				
GPS NMEA Sentences	Rx	Rx		Rx			
Communications functions also include Ethernet (CR310 only) and USB							

Digital I/O terminal functions									
	C1	C2	P_ SW	SE1	SE2	SE3	SE4	SE5	SE6
General I/O	✓	✓	✓	✓	✓	✓	✓		
Pulse-Width Modulation Output				<b>✓</b>	✓	<b>√</b>	<b>√</b>		
Interrupt	✓	✓		✓	✓	✓			



# **Global Sales & Support Network**

A worldwide network to help meet your needs



# Campbell Scientific regional offices

#### Australia

Location: Garbutt, QLD Australia Phone: 61.7.4401.7700

Email: info@campbellsci.com.au
Website: www.campbellsci.com.au

### Brazil

Location: São Paulo, SP Brazil Phone: 11.3732.3399

Email: vendas@campbellsci.com.br Website: www.campbellsci.com.br

### Canada

Location: Edmonton, AB Canada

Phone: 780.454.2505

Email: dataloggers@campbellsci.ca

Website: www.campbellsci.ca

### China

Location: Beijing, P. R. China Phone: 86.10.6561.0080

Email: info@campbellsci.com.cn Website: www.campbellsci.com.cn

#### Costa Rica

Location: San Pedro, Costa Rica
Phone: 506.2280.1564
Email: info@campbellsci.cc
Website: www.campbellsci.cc

#### France

Location: Vincennes, France
Phone: 0033.0.1.56.45.15.20
Email: info@campbellsci.fr
Website: www.campbellsci.fr

### Germany

Location: Bremen, Germany
Phone: 49.0.421.460974.0
Email: info@campbellsci.de
Website: www.campbellsci.de

### India

Location: New Delhi, DL India Phone: 91.11.46500481.482 Email: info@campbellsci.in Website: www.campbellsci.in

### South Africa

Location: Stellenbosch, South Africa

Phone: 27.21.8809960

Email: sales@campbellsci.co.za Website: www.campbellsci.co.za

### Spain

Location:Barcelona, SpainPhone:34.93.2323938Email:info@campbellsci.esWebsite:www.campbellsci.es

### **Thailand**

Location: Bangkok, Thailand Phone: 66.2.719.3399

Email: info@campbellsci.asia Website: www.campbellsci.asia

# UK

Location: Shepshed, Loughborough, UK

Phone: 44.0.1509.601141

Email: sales@campbellsci.co.uk
Website: www.campbellsci.co.uk

### USA

Location: Logan, UT USA *Phone:* 435.227.9120

Email: info@campbellsci.com
Website: www.campbellsci.com